



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0826
CALIBRATION DATE: 01-Oct-22

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999863e-001
rtca1 = 1.623128e-006
rtca2 = -2.916830e-008

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999880	0.9999879	-0.1
4.5000	0.9999930	0.9999930	0.0
15.0000	1.0000040	1.0000041	0.1
18.5000	1.0000060	1.0000064	0.4
24.0000	1.0000090	1.0000085	-0.5
29.0000	1.0000090	1.0000089	-0.1
32.5000	1.0000080	1.0000083	0.3

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

