



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0827
CALIBRATION DATE: 11-Sep-24

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.037651e+000
h = 1.590976e-001
i = -1.609300e-004
j = 4.015095e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -2.7313e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2555.81	0.00000	0.00000
0.9999	34.6404	2.96235	5011.07	2.96237	0.00002
4.5000	34.6192	3.26796	5198.46	3.26794	-0.00002
15.0000	34.5751	4.24514	5756.34	4.24512	-0.00002
18.5000	34.5657	4.58870	5939.85	4.58872	0.00002
24.0000	34.5558	5.14414	6224.84	5.14415	0.00001
29.0001	34.5505	5.66367	6479.82	5.66367	-0.00001
32.5000	34.5476	6.03441	6655.70	6.03448	0.00007

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

