

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0828
CALIBRATION DATE: 06-Nov-08

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999670e-001

rtca1 = 1.612646e-006

rtca2 = -2.899834e-008

BATH TEMP (ITS-90)	RTC FREO (Hz)	COMPUTED FREO (Hz)	RESIDUAL (PPM)
1.0000	0.9999690	0.9999686	-0.4
4.4999	0.9999730	0.9999737	0.7
15.0000	0.9999850	0.9999847	-0.3
18.4999	0.9999870	0.9999869	-0.1
23.9999	0.9999890	0.9999890	0.0
28.9999	0.9999890	0.9999894	0.4
32.5001	0.9999890	0.9999888	-0.2

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

