

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0828
CALIBRATION DATE: 07-Jul-10

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999669e-001

rtca1 = 1.662244e-006

rtca2 = -3.103341e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999690	0.9999685	-0.5
4.5000	0.9999730	0.9999737	0.7
15.0000	0.9999850	0.9999848	-0.2
18.5000	0.9999870	0.9999870	-0.0
24.0000	0.9999890	0.9999889	-0.1
29.0000	0.9999890	0.9999890	-0.0
32.5000	0.9999880	0.9999881	0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

