

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0828
CALIBRATION DATE: 07-Jul-10

SBE 37 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = -8.940484e-006
a1 = 2.860025e-004
a2 = -3.102494e-006
a3 = 1.757202e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
0.9999	573785.1	0.9999	0.0000
4.5000	491181.1	4.5000	-0.0000
15.0000	314105.0	15.0000	0.0000
18.5000	272274.5	18.5001	0.0001
24.0000	218785.5	23.9999	-0.0001
29.0000	180410.4	29.0000	-0.0000
32.5000	158143.4	32.5000	0.0000

Temperature ITS-90 = $1/[a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]] - 273.15$ (°C)

Residual = instrument temperature - bath temperature

