

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 0829
CALIBRATION DATE: 06-Nov-08

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.045615e+000

CPcor = -9.5700e-008

h = 1.579280e-001

CTcor = 3.2500e-006

i = -3.576541e-005

WBOTC = -2.8423e-005

j = 3.192678e-005

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2572.93	0.00000	0.00000
1.0000	34.7449	2.97044	5032.86	2.97045	0.00001
4.4999	34.7249	3.27694	5220.75	3.27693	-0.00001
15.0000	34.6812	4.25679	5780.02	4.25679	-0.00000
18.4999	34.6722	4.60130	5963.99	4.60129	-0.00001
23.9999	34.6621	5.15821	6249.73	5.15822	0.00002
28.9999	34.6563	5.67904	6505.35	5.67904	-0.00001
32.5001	34.6539	6.05088	6681.73	6.05088	-0.00000

$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

Conductivity = $(g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p)$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Residual = instrument conductivity - bath conductivity

