

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1242
CALIBRATION DATE: 16-Dec-09

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999611e-001

rtca1 = 1.739716e-006

rtca2 = -3.278579e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999630	0.9999628	-0.2
4.5000	0.9999680	0.9999683	0.3
15.0000	0.9999800	0.9999798	-0.2
18.5000	0.9999820	0.9999821	0.1
24.0000	0.9999840	0.9999840	-0.0
28.9999	0.9999840	0.9999840	-0.0
32.5000	0.9999830	0.9999830	0.0

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

