

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1244
CALIBRATION DATE: 12-Jun-09

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999843e-001

rtca1 = 1.803657e-006

rtca2 = -3.297446e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999860	0.9999861	0.1
4.4999	0.9999920	0.9999918	-0.2
15.0001	1.0000040	1.0000040	-0.0
18.5000	1.0000060	1.0000064	0.4
24.0001	1.0000090	1.0000086	-0.4
29.0000	1.0000090	1.0000089	-0.1
32.5000	1.0000080	1.0000081	0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

