



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1245
CALIBRATION DATE: 09-Sep-21

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999791e-001
rtca1 = 1.726217e-006
rtca2 = -3.122288e-008

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999810	0.9999808	-0.2
4.4999	0.9999860	0.9999862	0.2
15.0000	0.9999980	0.9999980	-0.0
18.5000	1.0000000	1.0000003	0.3
24.0000	1.0000030	1.0000025	-0.5
29.0001	1.0000030	1.0000029	-0.1
32.5000	1.0000020	1.0000022	0.2

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

