

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1246

CALIBRATION DATE: 16-Apr-12

SBE 37 PRESSURE CALIBRATION DATA

1500 psia S/N 193467

COEFFICIENTS:

PA0 = 1.082760e+000

PA1 = 6.010052e-002

PA2 = 2.499426e-008

PTCA0 = 1.461573e+002

PTCA1 = 8.089169e-001

PTCA2 = -1.122977e-003

PTCB0 = 3.012700e+001

PTCB1 = -4.000000e-004

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.58	383.2	23.0	14.25	-0.02
314.89	5372.1	23.0	314.86	-0.00
614.89	10320.9	22.8	614.28	-0.04
914.91	15258.6	22.7	914.25	-0.04
1214.98	20180.2	22.7	1214.45	-0.03
1515.02	25090.6	22.7	1515.18	0.01
1214.95	20194.4	22.7	1215.32	0.02
914.91	15276.2	22.7	915.32	0.03
614.91	10341.5	22.8	615.53	0.04
314.90	5381.0	22.7	315.41	0.03
14.58	389.3	22.3	14.65	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	407.10	0.00	30.13
29.00	404.30	30.00	30.11
24.00	400.56		
18.50	396.88		
15.00	393.58		
4.50	385.45		
1.00	382.78		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

16-Apr-12 0.00

