

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1251

CALIBRATION DATE: 28-Sep-15

SBE 37 PRESSURE CALIBRATION DATA

FSR: 1500 psia S/N 193472

COEFFICIENTS:

PA0 = 9.516254e-002

PA1 = 6.010107e-002

PA2 = 2.078541e-008

PTCA0 = -1.514375e+002

PTCA1 = 1.331373e-001

PTCA2 = -2.792294e-003

PTCB0 = 2.980200e+001

PTCB1 = -0.000000e+000

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.62	93.5	22.7	14.72	0.01
311.51	5007.7	23.0	310.62	-0.06
608.43	9921.0	23.0	607.47	-0.06
905.36	14822.9	23.0	904.63	-0.05
1202.31	19707.4	23.0	1201.73	-0.04
1499.18	24579.5	23.0	1499.07	-0.01
1202.30	19726.2	23.0	1202.88	0.04
905.38	14851.9	23.1	906.39	0.07
608.44	9956.5	23.1	609.62	0.08
311.47	5023.7	23.1	311.59	0.01
14.62	96.2	23.1	14.88	0.02

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT
32.50	105.90
29.00	105.18
24.00	105.24
18.50	105.57
15.00	105.92
4.50	104.68
1.00	104.04
TEMP (ITS90)	SPAN (mV)
0.00	29.80
30.00	29.80

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

