

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3531
 CALIBRATION DATE: 24-Jul-07

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 4979

COEFFICIENTS:

PA0 = -1.621966e-002
 PA1 = 2.409920e-002
 PA2 = 1.589910e-009

PTCA0 = 9.491881e+001
 PTCA1 = 7.626799e-002
 PTCA2 = -5.321921e-003
 PTCB0 = 2.512800e+001
 PTCB1 = -1.400000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.68	703.1	20.5	14.67	-0.00
107.40	4543.0	20.7	107.35	-0.01
207.30	8679.9	20.7	207.25	-0.01
307.34	12818.3	20.7	307.24	-0.02
407.25	16956.9	20.8	407.29	0.01
507.29	21091.3	20.7	507.28	-0.00
407.35	16959.3	20.7	407.34	-0.00
307.29	12822.8	20.7	307.34	0.01
207.31	8686.0	20.8	207.40	0.02
107.28	4541.3	20.7	107.31	0.01
14.68	703.9	20.8	14.69	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	732.68	-5.00	25.14
29.00	733.41	35.00	25.08
24.00	734.40		
18.50	735.37		
15.00	735.75		
4.50	735.99		
1.00	735.78		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

24-Jul-07 0.00

