

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3534
 CALIBRATION DATE: 28-Aug-06

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 5296

COEFFICIENTS:

PA0 = -6.070422e-002
 PA1 = 2.410315e-002
 PA2 = 1.785571e-009

PTCA0 = -6.160133e+002
 PTCA1 = 7.459370e-002
 PTCA2 = -8.849182e-003
 PTCB0 = 2.422388e+001
 PTCB1 = -2.250000e-004
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.58	-10.9	21.6	14.59	0.00
104.70	3724.8	21.7	104.68	-0.00
204.70	7868.6	21.7	204.68	-0.00
304.69	12009.3	21.7	304.66	-0.01
404.66	16147.7	21.7	404.64	-0.00
504.64	20283.8	21.7	504.63	-0.00
404.65	16148.8	21.8	404.67	0.00
304.66	12010.6	21.8	304.69	0.01
204.68	7869.8	21.8	204.71	0.00
104.70	3726.2	21.8	104.72	0.00
14.57	-11.8	21.8	14.57	-0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	26.80	-5.00	24.23
29.00	27.90	35.00	24.22
24.00	30.30		
18.50	31.80		
15.00	32.80		
4.50	33.60		
1.00	33.60		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

28-Aug-06 0.00

