

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3535
 CALIBRATION DATE: 28-Aug-06

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 5297

COEFFICIENTS:

PA0 = -7.341997e-002
 PA1 = 2.411739e-002
 PA2 = 1.893879e-009

PTCA0 = -5.665630e+002
 PTCA1 = 6.121522e-002
 PTCA2 = -9.434959e-003
 PTCB0 = 2.423275e+001
 PTCB1 = 1.500000e-004
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.58	36.7	21.5	14.55	-0.01
104.69	3773.5	21.6	104.69	0.00
204.62	7912.4	21.6	204.60	-0.00
304.64	12052.9	21.6	304.61	-0.01
404.62	16189.9	21.6	404.60	-0.00
504.60	20324.6	21.6	504.60	0.00
404.61	16191.2	21.7	404.63	0.01
304.63	12054.3	21.7	304.64	0.00
204.64	7914.5	21.7	204.65	0.00
104.66	3773.0	21.7	104.68	0.00
14.58	38.6	21.8	14.60	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	73.20	-5.00	24.23
29.00	74.60	35.00	24.24
24.00	77.60		
18.50	78.90		
15.00	79.70		
4.50	81.30		
1.00	81.10		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

28-Aug-06 -0.00

