

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3535

CALIBRATION DATE: 07-Sep-0

SBE 37 RTC CALIBRATION DATA

ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999902e-001

rtca1 = 1.690206e-006

rtca2 = -3.328945e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999920	0.9999919	-0.1
4.4999	0.9999970	0.9999971	0.1
15.0000	1.0000080	1.0000081	0.1
18.4999	1.0000100	1.0000101	0.1
23.9999	1.0000120	1.0000116	-0.4
29.0001	1.0000110	1.0000112	0.2
32.5001	1.0000100	1.0000100	-0.0

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

