

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 12120
CALIBRATION DATE: 29-Sep-15

SBE 37 V2 PRESSURE CALIBRATION DATA
FSR: 508 psia S/N 4134752

COEFFICIENTS:

PA0 = 7.735440e-002
PA1 = 1.580937e-003
PA2 = 5.534093e-012
PTempa0 = -8.234710e+001
PTempa1 = 4.834195e-002
PTempa2 = -3.745784e-007

PTCA0 = 5.250791e+005
PTCA1 = 4.755164e+000
PTCA2 = -1.150590e-001
PTCB0 = 2.510263e+001
PTCB1 = -2.750000e-004
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FS
14.58	534297.0	2225.0	14.58	0.00
104.84	591357.0	2226.0	104.83	-0.00
204.84	654555.0	2227.0	204.84	-0.00
304.84	717729.0	2227.0	304.85	0.00
404.84	780864.0	2228.0	404.85	0.00
504.85	843971.0	2228.0	504.84	-0.00
404.85	780871.0	2230.0	404.86	0.00
304.85	717733.0	2230.0	304.86	0.00
204.86	654561.0	2230.0	204.85	-0.00
104.85	591365.0	2230.0	104.85	-0.00
14.57	534299.0	2230.0	14.58	0.00

THERMAL CORRECTION

TEMP ITS90	THERMISTOR OUTPUT	INST OUTPUT
32.50	2421	534740.00
29.00	2346	534749.17
24.00	2239	534754.90
18.50	2121	534755.70
15.00	2046	534753.14
4.50	1822	534726.09
1.00	1748	534712.21
TEMP (ITS90)	SPAN (mV)	
-5.00	25.10	
35.00	25.09	

$y = \text{thermistor output}; t = \text{PTempa0} + \text{PTempa1} * y + \text{PTempa2} * y^2$

$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$

$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$

$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$

