

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4920
CALIBRATION DATE: 06-Mar-07

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999752e-001

rtca1 = 1.690112e-006

rtca2 = -3.075060e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999770	0.9999769	-0.1
4.5000	0.9999820	0.9999822	0.2
14.9999	0.9999940	0.9999937	-0.3
18.4999	0.9999960	0.9999960	-0.0
24.0000	0.9999980	0.9999981	0.1
28.9999	0.9999980	0.9999984	0.4
32.5000	0.9999980	0.9999977	-0.3

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

