

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4920
 CALIBRATION DATE: 04-Mar-09

SBE 37 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.019234e+000	CPcor = -9.5700e-008
h = 1.578953e-001	CTcor = 3.2500e-006
i = -6.677783e-005	WBOTC = 1.9888e-005
j = 3.260321e-005	

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2539.81	0.00000	0.00000
0.9999	34.8339	2.97732	5023.26	2.97732	0.00001
4.4999	34.8141	3.28453	5211.76	3.28453	-0.00000
15.0000	34.7710	4.26664	5772.38	4.26663	-0.00002
18.5000	34.7618	4.61192	5956.66	4.61191	-0.00001
24.0000	34.7514	5.17004	6242.73	5.17005	0.00001
29.0000	34.7452	5.69198	6498.53	5.69200	0.00002
32.5001	34.7413	6.06440	6674.84	6.06438	-0.00002

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

