

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4921
 CALIBRATION DATE: 28-Feb-07

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 10021

COEFFICIENTS:

PA0 = 5.271424e-002
 PA1 = 2.393267e-002
 PA2 = 2.782145e-009

PTCA0 = 2.541592e+001
 PTCA1 = 2.870960e-001
 PTCA2 = -8.749908e-003
 PTCB0 = 2.513688e+001
 PTCB1 = -8.250000e-004
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.50	631.0	20.6	14.50	0.00
104.78	4397.7	20.6	104.76	-0.00
204.76	8566.5	20.6	204.75	-0.00
304.75	12731.5	20.6	304.75	-0.00
404.73	16892.1	20.6	404.73	-0.00
504.71	21048.0	20.7	504.70	-0.00
404.73	16892.5	20.7	404.74	0.00
304.75	12731.8	20.8	304.76	0.00
204.76	8567.0	20.8	204.77	0.00
104.80	4399.0	20.8	104.80	-0.00
14.50	631.3	20.7	14.51	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	673.30	-5.00	25.14
29.00	674.30	35.00	25.11
24.00	675.20		
18.50	675.70		
15.00	675.60		
4.50	674.10		
1.00	673.80		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

● 28-Feb-07 -0.00

