

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4921
CALIBRATION DATE: 11-Mar-09

SBE 37 PRESSURE CALIBRATION DATA
508 psia S/N 10021

COEFFICIENTS:

PA0 = 3.052478e-002
PA1 = 2.393813e-002
PA2 = 2.649078e-009

PTCA0 = 2.733040e+001
PTCA1 = 1.830942e-001
PTCA2 = -5.997098e-003
PTCB0 = 2.513688e+001
PTCB1 = -8.250000e-004
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.82	646.7	19.8	14.84	0.00
105.08	4409.9	20.1	105.03	-0.01
205.08	8581.1	20.1	205.09	0.00
305.10	12747.0	20.0	305.11	0.00
405.09	16907.8	20.1	405.11	0.00
505.11	21064.2	20.0	505.09	-0.00
405.11	16907.9	20.0	405.11	0.00
305.11	12746.8	20.0	305.11	0.00
205.11	8581.3	20.0	205.09	-0.00
105.08	4411.7	20.1	105.07	-0.00
14.82	646.5	20.1	14.83	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	665.60	-5.00	25.14
29.00	666.30	35.00	25.11
24.00	667.15		
18.50	667.34		
15.00	667.38		
4.50	666.66		
1.00	666.30		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

11-Mar-09 0.00

