



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 9309
CALIBRATION DATE: 12-Sep-21

SBE 37 V2 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.167450e-004
a1 = 3.146440e-004
a2 = -5.276555e-006
a3 = 2.216938e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	579898.0	1.0001	0.0001
4.5001	495187.0	4.5000	-0.0001
15.0000	314496.0	15.0000	-0.0000
18.5000	272038.0	18.5002	0.0002
24.0000	217915.0	23.9999	-0.0001
29.0000	179218.0	28.9999	-0.0001
32.5000	156824.0	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

