



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2002
CALIBRATION DATE: 22-Sep-22

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999911e-001
rtca1 = 1.739721e-006
rtca2 = -3.278591e-008

BATH TEMPERATURE (° C)	RTC FREQUENCY (Hz)	COMPUTED FREQUENCY (Hz)	RESIDUAL (PPM)
1.0000	0.9999930	0.9999928	-0.2
4.5000	0.9999980	0.9999983	0.3
14.9999	1.0000100	1.0000098	-0.2
18.5000	1.0000120	1.0000121	0.1
24.0000	1.0000140	1.0000140	-0.0
29.0000	1.0000140	1.0000140	-0.0
32.5000	1.0000130	1.0000130	0.0

t = temperature (°C)

RTC frequency = $rtca0 + rtca1 * t + rtca2 * t^2$

Residual (PPM) = (Computed RTC frequency - Measured RTC frequency) * 1e6

