



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2448
CALIBRATION DATE: 15-Dec-17

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.031800e+000
h = 1.471652e-001
i = -2.749888e-004
j = 4.367885e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 7.2604e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2651.45	0.00000	0.00000
1.0000	34.7035	2.96724	5217.19	2.96724	-0.00000
4.5000	34.6837	3.27344	5412.53	3.27346	0.00001
15.0000	34.6418	4.25247	5993.64	4.25245	-0.00002
18.5000	34.6330	4.59667	6184.69	4.59666	-0.00000
24.0000	34.6233	5.15308	6481.31	5.15311	0.00003
29.0000	34.6175	5.67341	6746.50	5.67340	-0.00001
32.5000	34.6134	6.04460	6929.17	6.04434	-0.00026

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

