



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 2448
CALIBRATION DATE: 11-Oct-19

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.029829e+000
h = 1.468392e-001
i = -2.555291e-004
j = 4.276127e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 7.2604e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2651.46	0.00000	0.00000
1.0000	34.8295	2.97699	5226.68	2.97699	0.00000
4.5000	34.8087	3.28408	5422.50	3.28408	-0.00000
15.0000	34.7649	4.26597	6005.07	4.26597	-0.00000
18.5000	34.7553	4.61115	6196.56	4.61113	-0.00001
24.0000	34.7449	5.16918	6493.88	5.16920	0.00002
29.0000	34.7377	5.69089	6759.64	5.69089	0.00000
32.5000	34.7325	6.06303	6942.78	6.06302	-0.00001

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

