

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3318
 CALIBRATION DATE: 19-Jun-06

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 3683

COEFFICIENTS:

PA0 = 4.586750e-002
 PA1 = 2.399323e-002
 PA2 = 1.861956e-009

PTCA0 = 2.392846e+000
 PTCA1 = 2.090215e-001
 PTCA2 = -5.510080e-003
 PTCB0 = 2.502287e+001
 PTCB1 = -1.225000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.74	614.6	22.6	14.71	-0.01
99.85	4158.1	22.7	99.85	-0.00
199.84	8316.2	22.7	199.83	-0.00
299.83	12472.2	22.7	299.81	-0.00
404.81	16833.4	22.7	404.81	-0.00
504.80	20984.0	22.7	504.80	-0.00
404.80	16834.0	22.7	404.82	0.00
299.83	12472.9	22.7	299.83	0.00
199.84	8316.8	22.7	199.84	0.00
99.84	4158.5	22.8	99.86	0.00
14.74	617.2	22.8	14.77	0.01

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	648.31	-5.00	25.03
29.00	648.29	35.00	24.98
24.00	648.76		
18.50	649.22		
15.00	649.05		
4.50	648.07		
1.00	647.17		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

19-Jun-06 -0.00

