

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3318
 CALIBRATION DATE: 28-Feb-07

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 3683

COEFFICIENTS:

PA0 = 7.033894e-002
 PA1 = 2.400169e-002
 PA2 = 1.627946e-009

PTCA0 = 1.658316e+000
 PTCA1 = 2.904099e-001
 PTCA2 = -7.198673e-003
 PTCB0 = 2.502287e+001
 PTCB1 = -1.225000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.49	604.9	20.6	14.49	0.00
104.76	4360.6	20.6	104.76	-0.00
204.74	8518.3	20.7	204.74	-0.00
304.74	12674.2	20.7	304.73	-0.00
404.72	16827.5	20.8	404.72	0.00
504.70	20977.9	20.8	504.70	-0.00
404.72	16828.1	20.8	404.74	0.00
304.75	12675.1	20.8	304.76	0.00
204.76	8519.3	20.8	204.76	0.00
104.79	4361.5	20.9	104.78	-0.00
14.49	604.9	20.9	14.49	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	646.02	-5.00	25.03
29.00	646.69	35.00	24.98
24.00	647.16		
18.50	647.28		
15.00	646.91		
4.50	645.24		
1.00	644.71		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

● 28-Feb-07 -0.00

