

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3423
CALIBRATION DATE: 14-Apr-12

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999914e-001

rtca1 = 1.600614e-006

rtca2 = -3.279779e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999930	0.9999930	0.0
4.5000	0.9999980	0.9999980	-0.0
15.0000	1.0000080	1.0000081	0.1
18.4999	1.0000100	1.0000098	-0.2
24.0000	1.0000110	1.0000110	-0.0
29.0000	1.0000100	1.0000103	0.3
32.5000	1.0000090	1.0000088	-0.2

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

