

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3424
 CALIBRATION DATE: 31-Oct-08

SBE 37 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = -3.069615e-004
 a1 = 3.221852e-004
 a2 = -5.494645e-006
 a3 = 2.295764e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
0.9999	838113.0	0.9999	0.0000
4.5000	718949.7	4.5000	-0.0000
15.0000	462574.5	15.0000	-0.0000
18.5000	401773.1	18.5000	0.0000
24.0000	323847.7	24.0000	0.0000
29.0000	267796.1	28.9999	-0.0001
32.5000	235200.8	32.5001	0.0001

Temperature ITS-90 = $1/[a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]] - 273.15$ (°C)

Residual = instrument temperature - bath temperature

