



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Monterey Bay Aquarium Research Institute
Job Number 68578
Date 5/8/2012
Technician CJ

<u>Serial Number</u>	37SM34515-3424
----------------------	----------------

<u>Low Pressure (PSI)</u>	45	PSI
<u>Time (Minutes)</u>	15	Minutes

<u>High Pressure (PSI)</u>	10000	PSI
<u>Time (Minutes)</u>	30	Minutes

Pass ☒

Fail ☐

Comments

Replaced the main piston "O"-Rings.

