

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3429
CALIBRATION DATE: 11-Dec-07

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999850e-001

rtca1 = 1.612749e-006

rtca2 = -3.153742e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
1.0000	0.9999870	0.9999865	-0.5
4.5000	0.9999910	0.9999916	0.6
15.0000	1.0000020	1.0000021	0.1
18.5000	1.0000040	1.0000040	0.0
24.0000	1.0000060	1.0000055	-0.5
29.0000	1.0000050	1.0000052	0.2
32.5000	1.0000040	1.0000041	0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

