

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3531
CALIBRATION DATE: 01-Sep-0

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.845247e-001
h = 1.327990e-001
i = -1.028797e-004
j = 2.854077e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = -4.6806e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2723.64	0.00000	0.00000
1.0000	34.8803	2.98091	5458.56	2.98092	0.00001
4.4999	34.8600	3.28843	5665.39	3.28843	-0.00000
14.9999	34.8164	4.27161	6280.24	4.27160	-0.00002
18.5000	34.8070	4.61727	6482.27	4.61727	0.00000
24.0000	34.7966	5.17602	6795.82	5.17604	0.00002
29.0000	34.7910	5.69864	7076.15	5.69864	-0.00000
32.5000	34.7876	6.07155	7269.38	6.07155	-0.00000

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction

