

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3531
 CALIBRATION DATE: 28-Aug-06

SBE 37 PRESSURE CALIBRATION DATA
 508 psia S/N 4979

COEFFICIENTS:

PA0 = -1.632792e-002
 PA1 = 2.410643e-002
 PA2 = 1.260613e-009

PTCA0 = 9.279534e+001
 PTCA1 = 3.242104e-003
 PTCA2 = -1.573493e-003
 PTCB0 = 2.512800e+001
 PTCB1 = -1.400000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.57	696.5	21.8	14.57	-0.00
104.72	4430.2	21.9	104.71	-0.00
204.71	8569.2	21.9	204.68	-0.01
304.67	12706.6	21.8	304.64	-0.01
404.64	16843.1	21.8	404.64	-0.00
504.63	20977.3	21.8	504.61	-0.00
404.61	16843.8	21.8	404.65	0.01
304.65	12707.9	21.8	304.68	0.01
204.69	8570.0	21.9	204.70	0.00
104.73	4431.7	21.9	104.75	0.00
14.57	697.0	22.0	14.58	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	737.25	-5.00	25.14
29.00	737.63	35.00	25.08
24.00	737.94		
18.50	738.38		
15.00	738.42		
4.50	738.96		
1.00	738.71		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

28-Aug-06 -0.00

