

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3531
CALIBRATION DATE: 02-Dec-10

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999831e-001

rtca1 = 1.753212e-006

rtca2 = -3.434863e-008

BATH TEMP (ITS-90)	RTC FREQO (Hz)	COMPUTED FREQO (Hz)	RESIDUAL (PPM)
1.0000	0.9999850	0.9999848	-0.2
4.5000	0.9999900	0.9999903	0.3
15.0000	1.0000020	1.0000017	-0.3
18.5000	1.0000040	1.0000038	-0.2
24.0000	1.0000050	1.0000054	0.4
29.0000	1.0000050	1.0000051	0.1
32.5000	1.0000040	1.0000038	-0.2

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

