

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3531

CALIBRATION DATE: 28-Oct-13

SBE 37 PRESSURE CALIBRATION DATA

508 psia S/N 4979

COEFFICIENTS:

PA0 = -3.301755e-002

PA1 = 2.412569e-002

PA2 = 6.995686e-010

PTCA0 = 9.308773e+001

PTCA1 = 2.965108e-002

PTCA2 = -4.049184e-003

PTCB0 = 2.512800e+001

PTCB1 = -1.400000e-003

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.56	696.6	22.4	14.58	0.00
104.84	4431.4	22.3	104.81	-0.01
204.85	8570.8	22.2	204.83	-0.00
304.85	12709.0	22.2	304.86	0.00
404.86	16845.3	22.2	404.86	-0.00
504.86	20980.0	22.2	504.84	-0.00
404.86	16846.1	22.3	404.88	0.00
304.86	12710.1	22.3	304.88	0.00
204.86	8571.9	22.3	204.86	0.00
104.85	4432.1	22.3	104.82	-0.01
14.56	696.4	22.7	14.58	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	725.51	-5.00	25.14
29.00	726.22	35.00	25.08
24.00	726.90		
18.50	727.90		
15.00	728.53		
4.50	728.68		
1.00	728.82		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

28-Oct-13 0.00

