

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3722
CALIBRATION DATE: 03-Jun-08

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.041061e+000
h = 1.255645e-001
i = -6.677339e-005
j = 2.081903e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 8.4820e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2879.37	0.00000	0.00000
1.0000	35.2704	3.01105	5674.19	3.01107	0.00002
4.5000	35.2504	3.32161	5886.65	3.32158	-0.00003
15.0002	35.2070	4.31445	6518.87	4.31445	0.00000
18.5000	35.1979	4.66349	6726.72	4.66349	0.00001
24.0000	35.1874	5.22768	7049.42	5.22767	-0.00001
29.0001	35.1809	5.75528	7338.03	5.75529	0.00001
32.5000	35.1762	6.13160	7536.92	6.13159	-0.00001

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

