

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3722
CALIBRATION DATE: 01-Dec-10

SBE 37 PRESSURE CALIBRATION DATA
10153 psia S/N 6174

COEFFICIENTS:

PA0 = 8.326055e-001
PA1 = 4.916539e-001
PA2 = -4.895261e-007

PTCA0 = -2.033604e+001
PTCA1 = -5.185841e-002
PTCA2 = 5.988916e-003
PTCB0 = 2.458963e+001
PTCB1 = 9.250000e-004
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.65	10.4	21.7	15.10	0.00
2014.85	4096.3	21.7	2014.03	-0.01
4014.87	8218.6	21.7	4014.24	-0.01
6014.96	12376.3	21.8	6014.77	-0.00
8014.90	16568.4	21.8	8014.76	-0.00
10014.84	20795.7	21.8	10014.16	-0.01
8014.68	16570.4	21.8	8015.71	0.01
6014.72	12378.0	21.8	6015.58	0.01
4014.68	8219.9	21.8	4014.84	0.00
2014.73	4096.9	21.8	2014.31	-0.00
14.65	10.3	21.9	15.03	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	14.25	-5.00	24.59
29.00	13.28	35.00	24.62
24.00	11.95		
18.50	10.73		
15.00	10.18		
4.50	9.62		
1.00	9.61		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

