

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3722

CALIBRATION DATE: 13-Sep-14

SBE 37 RTC CALIBRATION DATA

ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999845e-001

rtca1 = 1.547121e-006

rtca2 = -3.277830e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999860	0.9999860	0.0
4.5000	0.9999910	0.9999908	-0.2
15.0000	1.0000000	1.0000004	0.4
18.4999	1.0000020	1.0000019	-0.1
24.0000	1.0000030	1.0000028	-0.2
29.0000	1.0000020	1.0000018	-0.2
32.5000	1.0000000	1.0000002	0.2

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

