



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3722
CALIBRATION DATE: 07-Sep-21

SBE 37 PRESSURE CALIBRATION DATA
10153 psia S/N 6174

COEFFICIENTS:

PA0 =	1.236935e+000	PTCA0 =	-1.918295e+001
PA1 =	4.920405e-001	PTCA1 =	-5.299897e-002
PA2 =	-4.904273e-007	PTCA2 =	5.760509e-003
		PTCB0 =	2.458963e+001
		PTCB1 =	9.250000e-004
		PTCB2 =	0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	TEMPERATURE (°C)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	INSTRUMENT OUTPUT (counts)
14.54	11.2	26.0	14.93	0.00	32.50	14.15
2001.70	4067.9	26.2	2000.84	-0.01	29.00	13.16
3988.92	8161.4	26.2	3988.46	-0.00	24.00	11.90
5976.19	12289.9	26.3	5976.42	0.00	18.50	10.80
7963.46	16452.4	26.3	7963.88	0.00	15.00	10.28
9951.14	20648.7	26.3	9950.32	-0.01	4.50	9.73
7963.38	16452.7	26.3	7964.03	0.01	1.00	9.76
5976.16	12290.6	26.4	5976.72	0.01		
3988.88	8162.2	26.4	3988.79	-0.00	TEMPERATURE (°C)	SPAN
2001.66	4068.4	26.4	2001.04	-0.01	-5.00	24.59
14.53	11.7	26.6	15.11	0.01	35.00	24.62

$$x = \text{instrument output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (PSIA)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)
● 07-Sep-21 0.00

