

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3777
 CALIBRATION DATE: 02-Jun-08

SBE 37 PRESSURE CALIBRATION DATA
 10153 psia S/N 6172

COEFFICIENTS:

PA0 = 2.775943e+000
 PA1 = 4.853371e-001
 PA2 = -4.894563e-007

PTCA0 = -1.346970e+001
 PTCA1 = 6.438449e-002
 PTCA2 = 8.790957e-003
 PTCB0 = 2.494788e+001
 PTCB1 = -6.250000e-004
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.66	17.1	22.1	14.84	0.00
2001.49	4123.2	22.3	2000.40	-0.01
3988.64	8267.9	22.3	3987.90	-0.01
5975.73	12449.1	22.3	5975.85	0.00
7962.85	16665.7	22.3	7963.28	0.00
9950.23	20916.8	22.3	9949.33	-0.01
7962.66	16666.1	22.3	7963.47	0.01
5975.59	12449.9	22.4	5976.22	0.01
3988.51	8269.3	22.4	3988.56	0.00
2001.39	4124.7	22.4	2001.10	-0.00
14.65	18.5	22.5	15.44	0.01

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	24.16	-5.00	24.95
29.00	22.14	35.00	24.93
24.00	19.48		
18.50	17.01		
15.00	15.71		
4.50	13.36		
1.00	12.87		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

02-Jun-08 0.00

