

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3777

CALIBRATION DATE: 17-Apr-12

SBE 37 PRESSURE CALIBRATION DATA

10153 psia S/N 6172

COEFFICIENTS:

PA0 = 2.829214e+000

PA1 = 4.853558e-001

PA2 = -4.896336e-007

PTCA0 = -1.210555e+001

PTCA1 = 4.953472e-002

PTCA2 = 8.859033e-003

PTCB0 = 2.494788e+001

PTCB1 = -6.250000e-004

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.66	19.0	23.0	15.11	0.00
2001.67	4124.7	22.9	2000.63	-0.01
3988.26	8268.4	23.0	3987.75	-0.01
5974.86	12448.2	22.9	5975.14	0.00
7961.61	16663.6	22.9	7962.10	0.00
9948.61	20913.5	22.9	9947.69	-0.01
7961.74	16664.4	22.9	7962.48	0.01
5974.70	12448.6	22.9	5975.33	0.01
3988.19	8268.9	22.8	3988.01	-0.00
2001.54	4125.4	22.8	2000.99	-0.01
14.66	19.3	22.8	15.30	0.01

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	24.67	-5.00	24.95
29.00	22.67	35.00	24.93
24.00	20.06		
18.50	17.65		
15.00	16.45		
4.50	14.19		
1.00	13.77		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

17-Apr-12 0.00

