



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3777
CALIBRATION DATE: 07-Sep-21

SBE 37 PRESSURE CALIBRATION DATA
10153 psia S/N 6172

COEFFICIENTS:

PA0 = 4.060091e+000
PA1 = 4.856264e-001
PA2 = -4.894363e-007

PTCA0 = -1.087300e+001
PTCA1 = 1.000792e-001
PTCA2 = 8.430574e-003
PTCB0 = 2.494788e+001
PTCB1 = -6.250000e-004
PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	TEMPERATURE (°C)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	INSTRUMENT OUTPUT (counts)
14.54	19.6	26.1	14.81	0.00	32.50	22.18
2001.70	4123.2	26.3	2000.54	-0.01	29.00	21.47
3988.92	8265.3	26.3	3988.23	-0.01	24.00	17.49
5976.19	12443.8	26.3	5976.34	0.00	18.50	15.21
7963.46	16657.6	26.4	7963.90	0.00	15.00	13.90
9951.14	20906.0	26.4	9950.18	-0.01	4.50	11.20
7963.38	16658.3	26.4	7964.23	0.01	1.00	10.68
5976.16	12444.9	26.4	5976.85	0.01	TEMPERATURE (°C)	SPAN
3988.88	8266.8	26.4	3988.93	0.00		
2001.66	4124.6	26.5	2001.18	-0.00		
14.53	21.0	26.6	15.36	0.01		
					-5.00	24.95
					35.00	24.93

$$x = \text{instrument output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (PSIA)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)
● 07-Sep-21 0.00

