

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3936
 CALIBRATION DATE: 13-Nov-06

SBE 37 PRESSURE CALIBRATION DATA
 10153 psia S/N 7325

COEFFICIENTS:

PA0 = 1.983494e+000
 PA1 = 4.883613e-001
 PA2 = -4.682396e-007

PTCA0 = -1.320813e+001
 PTCA1 = 3.120087e-001
 PTCA2 = -5.529649e-003
 PTCB0 = 2.482850e+001
 PTCB1 = 7.000000e-004
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.41	16.4	20.4	14.45	0.00
2001.27	4103.2	20.5	2001.22	-0.00
3988.20	8222.1	20.5	3987.78	-0.00
5975.15	12376.0	20.6	5975.15	0.00
7962.28	16564.3	20.6	7962.64	0.00
9949.77	20785.9	20.7	9949.30	-0.00
7962.40	16564.7	20.7	7962.80	0.00
5974.93	12376.2	20.8	5975.20	0.00
3987.86	8221.9	20.8	3987.65	-0.00
2000.77	4101.9	20.8	2000.56	-0.00
14.40	16.9	20.9	14.67	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	18.34	-5.00	24.82
29.00	18.50	35.00	24.85
24.00	18.40		
18.50	17.96		
15.00	17.47		
4.50	15.36		
1.00	14.39		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

13-Nov-06 -0.00

