

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3936
CALIBRATION DATE: 31-Oct-08

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999894e-001

rtca1 = 1.643851e-006

rtca2 = -3.305907e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999910	0.9999910	-0.0
4.5000	0.9999960	0.9999961	0.1
15.0000	1.0000070	1.0000066	-0.4
18.5000	1.0000080	1.0000085	0.5
24.0000	1.0000100	1.0000098	-0.2
29.0000	1.0000090	1.0000092	0.2
32.5000	1.0000080	1.0000079	-0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

