

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3936

CALIBRATION DATE: 12-Sep-14

SBE 37 PRESSURE CALIBRATION DATA

FSR: 10153 psia S/N 7325

COEFFICIENTS:

PA0 = 2.257747e+000

PA1 = 4.883885e-001

PA2 = -4.695172e-007

PTCA0 = -1.229075e+001

PTCA1 = 3.070371e-001

PTCA2 = -4.568122e-003

PTCB0 = 2.482850e+001

PTCB1 = 7.000000e-004

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.64	17.9	22.8	14.74	0.00
2014.73	4131.0	22.9	2014.19	-0.01
4014.84	8278.7	22.8	4014.43	-0.00
6015.10	12461.0	22.9	6014.98	-0.00
8014.95	16677.1	23.0	8015.08	0.00
10014.86	20927.0	22.9	10014.40	-0.00
8014.98	16678.1	23.0	8015.55	0.01
6015.21	12462.3	23.0	6015.57	0.00
4014.88	8280.0	22.9	4015.03	0.00
2014.86	4132.1	23.0	2014.72	-0.00
14.64	18.5	23.1	15.01	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT
32.50	19.53
29.00	19.42
24.00	19.15
18.50	18.50
15.00	17.91
4.50	15.70
1.00	14.66
TEMP (ITS90)	SPAN (mV)
-5.00	24.82
35.00	24.85

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

