



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 3936
CALIBRATION DATE: 22-Sep-22

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.020215e+000
h = 1.421108e-001
i = -2.374220e-004
j = 3.864785e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 7.1795e-006

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2682.54	0.00000	0.00000
1.0000	34.6741	2.96497	5298.75	2.96496	-0.00000
4.5000	34.6552	3.27102	5497.72	3.27102	0.00000
15.0000	34.6142	4.24944	6089.50	4.24943	-0.00001
18.5000	34.6056	4.59342	6284.03	4.59343	0.00001
24.0000	34.5957	5.14943	6585.98	5.14943	0.00000
29.0000	34.5891	5.66928	6855.93	5.66927	-0.00000
32.5000	34.5825	6.03981	7041.82	6.03982	0.00000

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

