

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3937
 CALIBRATION DATE: 06-Dec-07

SBE 37 PRESSURE CALIBRATION DATA
 10153 psia S/N 7328

COEFFICIENTS:

PA0 = 3.607883e+000
 PA1 = 4.952246e-001
 PA2 = -4.901903e-007

PTCA0 = 1.968216e+001
 PTCA1 = 4.949851e-001
 PTCA2 = -6.712365e-003
 PTCB0 = 2.455075e+001
 PTCB1 = 3.150000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.53	50.0	20.4	14.97	0.00
1999.53	4082.9	20.4	1998.93	-0.01
3999.71	8182.2	20.3	3999.36	-0.00
5999.91	12316.2	20.3	6000.08	0.00
7999.83	16483.9	20.3	8000.25	0.00
10000.68	20686.8	20.4	10000.01	-0.01
7999.89	16484.3	20.3	8000.44	0.01
5999.85	12316.4	20.3	6000.18	0.00
3999.72	8182.6	20.3	3999.55	-0.00
1999.59	4083.4	20.3	1999.21	-0.00
14.53	49.6	20.3	14.79	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	53.22	-5.00	24.54
29.00	53.00	35.00	24.66
24.00	52.30		
18.50	51.11		
15.00	50.11		
4.50	46.41		
1.00	44.71		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

06-Dec-07 -0.00

