

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3937
 CALIBRATION DATE: 11-Jun-09

SBE 37 PRESSURE CALIBRATION DATA
 10153 psia S/N 7328

COEFFICIENTS:

PA0 = 3.839546e+000
 PA1 = 4.952519e-001
 PA2 = -4.891949e-007

PTCA0 = 1.994014e+001
 PTCA1 = 5.129034e-001
 PTCA2 = -7.125286e-003
 PTCB0 = 2.455075e+001
 PTCB1 = 3.150000e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.60	50.3	21.7	14.99	0.00
1998.94	4082.4	21.8	1998.32	-0.01
3998.97	8181.9	21.8	3998.60	-0.00
6000.13	12317.7	21.8	6000.01	-0.00
7999.58	16485.0	21.9	7999.73	0.00
9999.74	20687.6	21.9	9999.36	-0.00
8000.27	16486.9	22.0	8000.53	0.00
5998.88	12317.0	22.0	5999.50	0.01
3999.22	8183.9	22.1	3999.39	0.00
1999.41	4084.3	22.1	1999.15	-0.00
14.60	49.9	22.1	14.76	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	53.15	-5.00	24.54
29.00	52.92	35.00	24.66
24.00	52.24		
18.50	51.06		
15.00	50.10		
4.50	46.20		
1.00	44.52		

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

11-Jun-09 -0.00

