

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4254
CALIBRATION DATE: 03-Jan-08

SBE 37 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.021347e+000
h = 1.514674e-001
i = -6.400989e-005
j = 3.019543e-005

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 8.9558e-006

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2596.15	0.00000	0.00000
1.0000	34.9738	2.98814	5137.06	2.98817	0.00003
4.5000	34.9540	3.29643	5329.97	3.29643	-0.00000
15.0000	34.9103	4.28192	5903.69	4.28184	-0.00008
18.5000	34.9008	4.62837	6092.32	4.62836	-0.00001
24.0000	34.8900	5.18837	6385.13	5.18843	0.00006
29.0000	34.8831	5.71203	6646.92	5.71208	0.00006
32.5000	34.8783	6.08558	6827.29	6.08552	-0.00006

$$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$$

$$\text{Conductivity} = (g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p) \text{ Siemens/meter}$$

$$t = \text{temperature}[^{\circ}\text{C}]; p = \text{pressure}[\text{decibars}]; \delta = \text{CTcor}; \epsilon = \text{CPcor};$$

$$\text{Residual} = \text{instrument conductivity} - \text{bath conductivity}$$

Date, Slope Correction

