

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4254
CALIBRATION DATE: 03-Jan-08

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999802e-001

rtca1 = 1.703704e-006

rtca2 = -3.485234e-008

BATH TEMP (ITS-90)	RTC FREO (Hz)	COMPUTED FREO (Hz)	RESIDUAL (PPM)
1.0000	0.9999820	0.9999819	-0.1
4.5000	0.9999870	0.9999872	0.2
15.0000	0.9999980	0.9999979	-0.1
18.5000	1.0000000	0.9999998	-0.2
24.0000	1.0000010	1.0000010	0.0
29.0000	1.0000000	1.0000003	0.3
32.5000	0.9999990	0.9999988	-0.2

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

