

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 4254
CALIBRATION DATE: 31-Oct-08

SBE 37 RTC CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

rtca0 = 9.999793e-001

rtca1 = 1.636612e-006

rtca2 = -3.073088e-008

BATH TEMP (ITS-90)	RTC FREQ (Hz)	COMPUTED FREQ (Hz)	RESIDUAL (PPM)
0.9999	0.9999810	0.9999809	-0.1
4.5000	0.9999860	0.9999861	0.1
15.0000	0.9999970	0.9999970	-0.0
18.5000	0.9999990	0.9999991	0.1
24.0000	1.0000010	1.0000009	-0.1
29.0000	1.0000010	1.0000010	-0.0
32.5000	1.0000000	1.0000001	0.1

$$\text{RTC frequency} = \text{rtca0} + \text{rtca1} * t + \text{rtca2} * t^2$$

$$\text{Residual} = (\text{Computed RTC frequency} - \text{Measured RTC frequency}) * 1e6$$

Date, Delta F ppm

